



Study Title

Antimicrobial Activity and Efficacy of VYV's Device

Test Method

Custom Device Study Based on: ASTM E1153

Study Identification Number

NG17374

Study Sponsor

Edward Kiegle
VYV
kiegle@vyv.com

Test Facility

Microchem Laboratory
1304 W. Industrial Blvd
Round Rock, TX 78681
(512) 310-8378

Report Author: Samuel Hanley, B.S.

Purpose of the Study

The purpose of this study was to determine the antimicrobial efficacy of VYV's submitted test device.

Brief History of the Performing Laboratory

Microchem Laboratory is located in the greater Austin, Texas area. It is owned and operated by microbiologist Dr. Benjamin Tanner. The core of the company was founded by Dr. Tanner as Antimicrobial Test Laboratories in 2006. Antimicrobial Test Laboratories was later combined with a niche cosmetic testing lab and Microchem Laboratory, founded in 1988 by Dr. Norman Miner. The combined labs have operated under one roof as Microchem Laboratory since 2016. Microchem Laboratory is ISO 17025 accredited and offers testing in compliance with current Good Laboratory Practice (GLP) regulations as stipulated by EPA and FDA. Clients are always welcome to tour the lab, observe studies, and audit the lab's quality systems.

Study Timeline

Devices Received	Cultures Initiated	Carriers Inoculated	Carriers Treated	Enumeration Plates Evaluated	Report Delivered
05FEB2021	25MAR2021	25MAR2021	25MAR2021	26MAR2021	31MAR2021

Amended report sent on 05APR2021


Test Device Information

Name of Test Device: VYV Benchtop Unit
Manufacturer: VYV
Mode of Action: Germicidal Visible Light (405 nm)

A description of how to operate the device was provided by the Study Sponsor prior to test initiation.

Test Microorganism Information

The test microorganism(s) selected for this test:




MS2 Bacteriophage (MS2), ATCC 15597-B1

This virus is a non-enveloped positive-stranded RNA virus of the bacteriophage family Leviviridae. Bacterial cells are the hosts for bacteriophages, and *E. coli* 15597 serves this purpose for MS2 bacteriophage. Its small size, icosahedral structure, and environmental resistance has made MS2 ideal for use as a surrogate virus (particularly in place of picornaviruses such as poliovirus and human norovirus) in water quality and disinfectant studies.

Permissive Host Cell System for MS2: *Escherichia coli*, 15597

Diagram of the Test Procedure



Summary of the Procedure

- Test microorganism is prepared in appropriate liquid broth.
- Test microorganism is harvested and the resulting suspension is diluted to achieve $\geq 1 \times 10^6$ PFU/mL.
- Test and control carriers are inoculated and allowed to dry in optimal conditions for test microorganism.
- Test carriers are placed in test device for the Sponsor-determined contact time.
- Test carriers are harvested into liquid media and plated in optimal incubation conditions and time for the test microorganism.
- After incubation, microbial concentrations are determined and reductions relative to pre-treatment controls are calculated.

Criteria for Scientific Defensibility of a Custom Device Study

For Microchem Laboratory to consider a Device Study study to be scientifically defensible, the following criteria must be met:

1. The initial and final concentration of microorganisms must be significantly high enough to observe the passing criteria/log reduction.
2. The media used for testing must be sterile.
3. The target microorganism must be pure colony morphology.

Passing Criteria

Due to the modified nature of the study, passing criteria may be determined by the Study Sponsor prior to test initiation. If no passing criteria is established, a conclusion about the data is not provided by Microchem Laboratory, but the Study Sponsor may determine significance based on statistical interpretation or other means.

Testing Parameters

Culture Growth Media:	N/A, Freezer Stock	Culture Growth Time:	N/A
Culture Dilution Media	Phosphate Buffered Saline, Artificial Saliva	Culture Supplement	N/A
Carrier Type	1" x 3" Glass Slides	Inoculum Volume	0.020 ml
Carrier Dry Time	30 minutes	Carrier Dry Temp. and Humidity	23.1°C / 43%
Contact Time	6 Hours	Contact Temperature	Ambient
Irradiance	2 mW/cm ²	Harvest Media (Volume)	Phosphate Buffered Saline w/ 0.1% Triton X-100 (20 ml)
Enumeration Media	50% Tryptic Soy Broth	Incubation Temperature	36°C
Incubation Time	18-24 Hours		

Study Notes

Per study sponsor request artificial saliva was used to dilute the inoculum. The artificial saliva was prepared using study sponsor provided instructions.

Amended report on 05APR2021 to include updated photo of the wattage when the device is running, changed mode of action from UV light (Germicidal) to Germicidal Visible light (405nm), and included the irradiance of the device in the test parameters (this information was provided by the study sponsor and not taken at Microchem).

Study Photographs




Image 1: Testing set up. Inoculum diluted in PBS is on the left and the inoculum diluted in the artificial saliva is on the right.

Study Photographs



Image 2: Kill-A-WATT reading when device is on.

Control Results

Neutralization Method: N/A

Media Sterility: No Growth

Growth Confirmation: Pure and Viable

Calculations

PFU/ml = (Average plate count) x 1:10 serial dilution factor

PFU/carrier = (Average plate count) x 1:10 serial dilution factor x media dilution factor

PFU/carrier = PFU/ml x total harvest media volume

Percent Reduction = $\frac{(B - A)}{B} \times 100\%$

$\text{Log}_{10} \text{ Reduction} = \text{Log}(B/A)$


Where:

B = Number of viable test microorganisms on the control carriers immediately after inoculation

A = Number of viable test microorganisms on the test carriers after the contact time

Results of the Study

Test Microorganism	Device	Culture Diluent	Contact Time	Contact Distance	Replicate	PFU/Carrier	Average PFU/Carrier	Percent Reduction Compared to Parallel Control	Log ₁₀ Reduction Compared to Parallel Control
MS2 Bacteriophage ATCC 15597-B1	Control	PBS	Time Zero	N/A	1	3.60E+06	1.73E+06	N/A	
					2	2.90E+06			
					3	3.70E+06			
			Time Final		1	7.90E+04	6.67E+04		
					2	4.90E+04			
					3	7.20E+04			
	VYV Benchtop Unit	PBS	6 Hours	6.5"	1	< 1.00E+01	< 1.00E+01	> 99.985%	> 3.82
					2	< 1.00E+01			
					3	< 1.00E+01			
	Control	Artificial Saliva	Time Zero	N/A	1	3.90E+05	5.60E+05	N/A	
					2	5.10E+05			
					3	7.80E+05			
			Time Final		1	6.40E+04	8.00E+04		
					2	9.80E+04			
					3	7.80E+04			
	VYV Benchtop Unit	Artificial Saliva	6 Hours	6.5"	1	2.70E+02	4.70E+02	99.41%	2.23
					2	9.80E+02			
					3	1.60E+02			
Note: The lower limit of detection for this study was 1.00E+01 PFU/carrier. Values observed less than the limit are reported as "<1.00E+01" in the results table and zero in the graph.									



The results of this study apply to the tested substances(s) only. Extrapolation of findings to related materials is the responsibility of the Sponsor.

Copyright © Microchem Laboratory, 2021. Reproduction and ordinary use of this study report by the entity listed as "Sponsor" is permitted. Other copying and reproduction of all or part of this document by other entities is expressly prohibited, unless prior permission is granted in writing by Microchem Laboratory.

